


ultra² Buffer B Reagent

REF 01-03-0011	Fill 940mL
REF 01-03-0012	Fill 3.8L



Trinity Biotech
 Kansas City, MO 64132 USA
 Tel: 1 800-325-3424
 Fax: 1 816-361-1974



Trinity Biotech plc
 Bray Co. Wicklow, Ireland
 Tel. 353 1 2769800
 Fax 353 1 2769888
 www.trinitybiotech.com

Rev D
 03/17

Pour d'autres langues	Para outras línguas
Für andere Sprachen	Για τις άλλεςλώσσες
Para otras lenguas	För andra språk
Per le altre lingue	For andre språk
Dla innych języków	



www.trinitybiotech.com


INTENDED USE


Buffer B Reagent is designed for use with the Trinity Biotech ultra2 systems for the quantitative measurement of the percentage of total glycosylated hemoglobins in whole blood. For *in vitro* diagnostic use only. **IVD**

SUMMARY AND EXPLANATION OF TEST

A_{1c}/GHb - Assessment of glycosylated hemoglobin has proven useful in the control of diabetes. Reagents are performance validated to assure accuracy and precision with the Primus assay and system for the measurement of glycosylated hemoglobin. Reagent is ready for use.

STORAGE AND STABILITY

 Store at ambient temperatures (2 – 28°C). Do not allow to freeze. The reagent is stable until the expiration date indicated on the label when kept tightly closed and protected from extreme environmental conditions.

 See the container label for the expiration date. **DO NOT USE** after the expiration date.

PRECAUTIONS

For *in vitro* diagnostic use only. Avoid skin contact. Consult the product SDS for safety information. This reagent is used in conjunction with blood testing equipment and warrants handling under universal precaution procedures for safety.

ORDERING INFORMATION

Reference No.	Item	Quantity
01-03-0011	Buffer B	940mL
01-03-0012	Buffer B	3.8L



Manufacturer



European Conformity



Authorized Representative



Reference Number



Lot Number



For *in vitro* Diagnostic Use



Expiration



Temperature limitation



Consult Instructions for Use