

G-6-PDH CONTROL, DEFICIENT

REF G5888

6 x 0.5 mL

Pour d'autres langues
Für andere Sprachen
Para otras lenguas
Per le altre lingue
Dla innych języków

Para outras línguas
Για τις άλλεςλώσεις
För andra språk
For andre språk
For andre sprog


www.trinitybiotech.com

ASSIGNED VALUES

The assigned values are lot specific and can be found on the enclosed component list.

INTENDED USE

G-6-PDH Control, Deficient, is designed for use in the quality control of Trinity Biotech's procedures which determine glucose-6-phosphate dehydrogenase (G-6-PDH) quantitatively using Trinity Biotech Procedure No. 345-UV or qualitatively using Trinity Biotech Procedure Nos. 203 or 400. The G-6-PDH Control serves to verify procedure reliability when assayed with test samples.

REAGENT

G-6-PDH Control, Deficient, Catalog No. G 5888
Lyophilized G-6-PDH (microbial) in a human red cell hemolysate base with stabilizers and preservatives.

PRECAUTIONS

G-6-PDH Control is for "in vitro Diagnostic Use". Normal precautions exercised in handling laboratory reagents should be followed. Dispose of waste observing all local, state and federal laws.

G-6-PDH Control is a POTENTIALLY BIOHAZARDOUS MATERIAL. Source materials from which this product was derived were found negative for HBsAg and for antibodies against HCV, HIV-1 and HIV-2 by approved test methods. However, since no test method can offer complete assurance that infectious agents are absent, this product should be handled observing the same safety precautions employed when handling any potentially infectious material.

Refer to Material Safety Data Sheet for any updated risk, hazard or safety information.

PREPARATION

1. Add 0.5 ml water to vial of G-6-PDH Control.
2. Allow to stand 5 minutes and then swirl gently.
3. Swirl intermittently until dissolution is complete.

STORAGE AND STABILITY

Store unopened vials in the refrigerator (2–8°C). Vial label bears expiration date.

After reconstitution, G-6-PDH Control solution is stable for 1 week refrigerated (2–8°C) or 2 weeks frozen (–20°C) with a maximum of three freeze-thaw cycles.

DETERIORATION

G-6-PDH Control solution should be discarded if turbidity develops.












GENERAL INSTRUCTIONS FOR USE

After reconstitution, G-6-PDH Control Solution is ready for use and should be treated in exactly the same manner as a test blood sample. G-6-PDH Control solution should be swirled gently to mix before each use.

Trinity Biotech, Inc., warrants that its products conform to the information contained in this and other Trinity Biotech publications. Purchaser must determine the suitability of the product for its particular use.

ORDERING INFORMATION

KIT	G-6-PDH, Deficient	
Catalog No.	Item	Quantity
G5888	G-6-PDH Controls	
G5029	Deficient	6x0.5 mL
G6888	Intermediate	6x0.5 mL
	Normal	6x0.5 mL

 Manufactured	CONTROL N Normal Control	CONT Contents
EC REP Authorized Representative	CONTROL I Intermediate Control	SUBS Substrate
 Consult accompanying documents	CONTROL D Deficient Control	REAG Reagent
REF Product Number	CAL Calibrator	DIL Diluent
LOT Lot	C.F. Coefficient Factor	ACT Activator
 Use by	RNG Range	IVD For In Vitro Diagnostic use
 Caution, consult accompanying documents	STD Standard	 Biohazard
 Store at 2–8°C	 Store at 2–30°C	 or  Hazard
 Trinity Biotech USA Jamestown, NY 14701 Tel. 1 800-325-3424 Fax: 716-488-1990	EC REP 	Trinity Biotech plc Bray Co. Wicklow, Ireland Tel. 353 1 2769800 Fax 353 1 2769888 www.trinitybiotech.com

G5888-29 Rev E
5/2013